



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 & ANSI/NCSL Z540-1-1994

Leeds Precision Instruments, Inc.

17300 Medina Road
Suite 600
Plymouth, MN 55447-5639
Peter McCullar Phone: 763-546-8575

CALIBRATION

Valid to: October 16, 2013

Certificate Number: AC-1295

I. Dimensional

Table with 5 columns: PARAMETER / EQUIPMENT, RANGE, CALIBRATION AND MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(±)], REFERENCE STANDARD OR EQUIPMENT, METHOD(S). Row 1: Microscopes, Up to 12 in / Up to 305 mm, 578 μin / 14.5 μm, Glass Scale / Stage Micrometer, PR-0001

Notes:

- 1. Calibration and Measurement Capabilities (Expanded Uncertainties) are based on approximately a 95% confidence interval, using a coverage of k=2
2. On-site calibration service is available
3. This scope is part of and must be included with the Certificate of Accreditation No. AC-1295

Handwritten signature of Keith Greenway

Vice President

